



# Analytical Test Report

Science First. ORELAP ID: 4096 OLCC#: 010-1002112892C

7200 Johnson Creek Blvd., Portland, OR 97206 (503) 307-0096

## Hand Sanitizer

Lab ID: 2210171-02

Myxner Systemz

METRC Batch ID: B#22HS10

Date Sampled: 10/31/22

Date Printed: 11/4/22

Report cannot be used for OLCC/OHA compliance.

## Potency Analysis

Analytical Method: De Backer, Journal of Chromatography b.2009. 11.004 -  
SOP 102 - Cannabinoids via High Performance Liquid Chromatography

Cannabinoids	mg/g	LOQ	Notes
THCA	< LOQ	0.0491	
delta 9-THC	< LOQ	0.0491	
delta 8-THC	< LOQ	0.0491	
CBDA	< LOQ	0.0491	
CBDVA	< LOQ	0.0491	
CBGA	< LOQ	0.0491	
CBD	3.98	0.0491	
CBDV	< LOQ	0.0491	
CBG	2.05	0.0491	
CBN	< LOQ	0.0491	
CBC	< LOQ	0.0491	
Total CBG	2.05	0.0491	
Total Cannabinoids	6.04	0.0491	

**Total THC**  
**< LOQ** mg/g

**Total CBD**  
**3.98** mg/g

&lt;LOQ - Results below the Limit of Quantitation

Acid form of THC/CBD are decarboxylated by heat, lose 12% of original mass as CO<sub>2</sub>. Result = "bioactive"

"Total" Cannabinoid accounts for decarboxylation and moisture content. Total THC = [(THCA×0.877) + Δ9THC] / (100%-MC)

**Chris Griffey**  
Lab Director

Sample tested in compliance with OAR 333-007 (TNI standards). Test results meet all requirements of NELAP and the Rose City Laboratories quality assurance plan unless otherwise noted. Samplings performed by RCL personnel follow sampling SOP 30. Samples not collected by RCL personnel are analyzed in "as received" condition. These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of Rose City Labs.



# Analytical Test Report

Science First. ORELAP ID: 4096 OLCC#: 010-1002112892C

7200 Johnson Creek Blvd., Portland, OR 97206 (503) 307-0096

## Hand Sanitizer

Myxner Systemz

Laboratory ID: 2210171-02

## Quality Control Potency

Batch: B22K020 - Potency

### Blank(B22K020-BLK1)

Analyte	Result	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	< LOQ	0.0500	mg/g		11/03/22 21:45	11/04/22 01:34	
delta 9-THC	< LOQ	0.0500	mg/g		11/03/22 21:45	11/04/22 01:34	
CBGA	< LOQ	0.0500	mg/g		11/03/22 21:45	11/04/22 01:34	
CBDA	< LOQ	0.0500	mg/g		11/03/22 21:45	11/04/22 01:34	
CBD	< LOQ	0.0500	mg/g		11/03/22 21:45	11/04/22 01:34	
CBN	< LOQ	0.0500	mg/g		11/03/22 21:45	11/04/22 01:34	
CBG	< LOQ	0.0500	mg/g		11/03/22 21:45	11/04/22 01:34	
delta 8-THC	< LOQ	0.0500	mg/g		11/03/22 21:45	11/04/22 01:34	
CBC	< LOQ	0.0500	mg/g		11/03/22 21:45	11/04/22 01:34	
Total CBG	< LOQ	0.0500	mg/g		11/03/22 21:45	11/04/22 01:34	
CBDVA	< LOQ	0.0500	mg/g		11/03/22 21:45	11/04/22 01:34	
CBDV	< LOQ	0.0500	mg/g		11/03/22 21:45	11/04/22 01:34	

### LCS(B22K020-BS1)

Analyte	% Recovery	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	98.5	0.0512	mg/g	90-110	11/03/22 21:45	11/04/22 01:50	
delta 9-THC	105	0.0512	mg/g	90-110	11/03/22 21:45	11/04/22 01:50	
CBDA	96.6	0.0512	mg/g	90-110	11/03/22 21:45	11/04/22 01:50	
CBD	104	0.0512	mg/g	90-110	11/03/22 21:45	11/04/22 01:50	
delta 8-THC	101	0.0512	mg/g	90-110	11/03/22 21:45	11/04/22 01:50	

**Chris Griffey**  
Lab Director



# Analytical Test Report

Science First. ORELAP ID: 4096 OLCC#: 010-1002112892C

7200 Johnson Creek Blvd., Portland, OR 97206 (503) 307-0096

## Hand Sanitizer

*Myxner Systemz*

Laboratory ID: 2210171-02

### Notes and Definitions

- B Analyte detected in method blank, but not associated samples.
  - B2 Analyte detected in sample and associate method blank.
  - C Interference due to co-elution.
  - D Initial result exceeded calibration range, reported data are based on analysis of a dilution.
  - H Non-homogenous sample matrix affecting RPD and/or QC.
  - I Manual Integration was performed.
  - L Duplicate sample relative percent difference (RPD) exceeds QC limits.
  - M Anomalous results due to matrix interference
  - P Peaks manually split.
  - Q1 QC out of limits but still ok
  - Q2 Quality Control outside QC limits. Data considered estimate.
  - Q3 CCV was above the acceptance criteria. Non-detect samples are considered acceptable.
  - Q4 CCV was below the acceptance criteria, however the sample still exceeds the regulatory limit.
  - R Marginal Exceedence.
  - U Reported result is an estimate. The analyte was detected above the calibration range.
  - X Problems with initial analysis, reported data are from reinjection of prepared sample.
- <LOQ - Results below the Limit of Quantitation - Compound not detected

A handwritten signature in black ink that reads "Chris Griffey".

**Chris Griffey**  
Lab Director